

Reagents also compatible in Hach®

VARIO Powder Packs (PP) and Reagents

Test	Range	Reagent	Liquid Reagent	Tube Tests	Powder Pack
Aluminium	0 – 0.22 mg/l Al	VARIO Aluminium Reagent, Set F20 consists of: VARIO Aluminium ECR VARIO Aluminium Hexamine VARIO Aluminium Masking Rgt	■		■ ■
Ammonia	0 – 0.5 mg/l N	VARIO Ammonia Nitrogen, Set F10 consists of: VARIO Ammonia Salicylate, F10 VARIO Ammonia Cyanurate, F10			■ ■
Ammonia LR	0 – 2.5 mg/l N	VARIO Am tube test Reagent, Set LR, F5 consists of: VARIO Ammonia Salicylate, F5 VARIO Ammonia Cyanurate, F5 VARIO Am Diluent Reagent Low Range		■	■ ■
Ammonia HR	0 – 50 mg/l N	VARIO Am tube test Reagent, Set HR, F5 consists of: VARIO Ammonia Salicylate, F5 VARIO Ammonia Cyanurate, F5 VARIO Am Diluent Reagent High Range		■	■ ■
Bromine	0.05 – 4.5 mg/l Br	VARIO Chlorine TOTAL-DPD, F10 VARIO Chlorine TOTAL-DPD, F10			■ ■
Chlorine free, combined and total Chlorine dioxide	Visual Test Kit up to 3.5mg/l Cl ₂	VARIO Chlorine FREE-DPD, F5 VARIO Chlorine FREE-DPD, F5 VARIO Chlorine TOTAL-DPD, F5 VARIO Chlorine TOTAL-DPD, F5 VARIO Chlorine FREE-DPD, F10 VARIO Chlorine FREE-DPD, F10 VARIO Chlorine TOTAL-DPD, F10 VARIO Chlorine TOTAL-DPD, F10 VARIO Chlorine FREE-DPD, F25 VARIO Chlorine FREE-DPD, F25 VARIO Chlorine TOTAL-DPD, F25 VARIO Chlorine TOTAL-DPD, F25			■ ■ ■ ■ ■ ■ ■ ■ ■ ■ ■ ■
	0.01 – 2 mg/l Cl ₂				
	0 – 5 mg/l Cl ₂				
Chlorine, online free and total	0,035 - 5 mg/l Cl ₂	Chlorine FREE, Set consists of: Chlorine, DPD Compound (free & total) Chlorine FREE, Indicator Solution Chlorine FREE, Buffer Solution	■ ■		■
		Chlorine TOTAL, Set consists of: Chlorine, DPD Compound (free & total) Chlorine TOTAL, Indicator Solution Chlorine TOTAL, Buffer Solution	■ ■		■
COD HR	0 – 15000 mg/l O ₂	COD VARIO 0 - 15000 mg/l		■ ■ ■	
Copper	0 – 5 mg/l Cu	VARIO CU1, F10 VARIO CU1, F10			■ ■
DEHA	20 - 500 µg/l DEHA	VARIO DEHA REAGENT SET consists of: VARIO OXYSCAV 1 RGT VARIO DEHA 2 RGT	■		■

instruments*

Method	Applications	Quantity	Code
Eriochrome cyanine R	Water	1 Set	53 50 00
		100 100 25 ml	
Salicylate	Water, waste water, seawater	1 Set	53 55 00
		2 x 100 2 x 100	
Salicylate	Water, waste water, seawater	1 Set	53 56 00
		50 50 50 tubes	
Salicylate	Water, waste water, seawater	1 Set	53 56 50
		50 50 50 tubes	
DPD-Method: USEPA accepted for drinking water analysis	Water, waste water, seawater	100	53 01 90
		1000	53 01 93
DPD method: USEPA accepted for drinking water analysis	Water, waste water, seawater	100	53 00 90
		1000	53 00 93
		100	53 00 80
		1000	53 00 83
DPD method: USEPA accepted for drinking water analysis	Water, waste water, seawater	100	53 01 80
		1000	53 01 83
		100	53 01 90
		1000	53 01 93
DPD method: USEPA accepted for drinking water analysis	Water, waste water, seawater	100	53 01 10
		1000	53 01 13
		100	53 01 30
		1000	53 01 33
DPD-method: USEPA accepted for drinking water analysis	for use in Hach® CL17 Process Analysers	1 Set	53 02 10
		24 g	53 02 00
		473 ml	53 02 22
		473 ml	53 02 23
DPD-method: USEPA accepted for drinking water analysis	for use in Hach® CL17 Process Analysers	1 Set	54 02 10
		24 g	53 02 00
		473 ml	54 02 22
		473 ml	54 02 23
Dichromate Reactor, Digestion	Water, waste water, seawater	25 tubes	2 42 07 22
		150 tubes	2 42 07 27
		25 tubes, mercury free	2 42 07 12
		150 tubes, mercury free	2 42 07 16
Bicinchoninate	Water, waste water, seawater	100	53 03 00
		1000	53 03 03
PPST		1 Set	53 60 00
		100 100 ml	



MSDS (Material Safety Data Sheets): www.lovibond.com

* HACH® is a registered trademark of Hach Company, Loveland, Colorado. The use of the HACH® trademark does not imply any affiliation with or approval by Hach Company regarding the formulation, testing or compatibility of these products for use in HACH® brand spectrophotometers or other devices or systems.

Reagents also compatible in Hach®

VARIO Powder Packs (PP) and Reagents

Test	Range	Reagent	Liquid Reagent	Tube Tests	Powder Pack
Hydrazine	0.005 – 0.6 mg/l N ₂ H ₄	VARIO Hydra2 Reagent	■		
Iron (Fe ²⁺ , Fe ³⁺), dissolved	0 – 3 mg/l Fe 0 – 1.8 mg/l Fe	VARIO Ferro, F10 VARIO IRON TPTZ			■ ■
Iron, total, Fe in Mo	0.01 - 1.8 mg/l	VARIO (Fe in Mo) Reagent Set consists of: VARIO (Fe in Mo) Rgt 1 VARIO (Fe in Mo) Rgt 2			■ ■
Manganese LR	0 – 0.7 mg/l Mn	VARIO Manganese Reagent, Set LR, F10 consists of: VARIO Alkaline-Cyanide Reagent Solution VARIO Ascorbic Acid VARIO PAN Indicator Solution	■ ■		■
Manganese HR	0 – 20 mg/l Mn	VARIO Manganese Reagent, Set HR, F10 consists of: VARIO MANGANESE CITRATE BUFFER, F10 VARIO SODIUMPERIODATE, F10			■ ■
Molybdate LR	0.5 – 5 mg/l MoO ₄	VARIO MOLYBDENUM LR, Set, F10 consists of: VARIO Molybdenum 1 LR, F10 VARIO Molybdenum 2 LR, F10			■ ■
Molybdate HR	0 – 35 mg/l Mo	VARIO MOLYBDENUM HR, Set F10 consists of: VARIO MOLYBDENUM HR1, F10 VARIO MOLYBDENUM HR2, F10 VARIO MOLYBDENUM HR3, F10			■ ■ ■
	0 – 35 mg/l Mo	VARIO MOLYBDENUM HR, Set F25 consists of: VARIO MOLYBDENUM HR1, F25 VARIO MOLYBDENUM HR2, F25 VARIO MOLYBDENUM HR3, F25			■ ■ ■
Nitrate	0 – 30 mg/l N	VARIO NITRA X Reagent, Set consists of: VARIO NITRA X Test vials VARIO NITRA NITROGEN NITRATE Reag. B Deionised water		■	■
Nitrogen, total LR	0 – 25 mg/l N	VARIO TOTAL NITROGEN LR, Set consists of a) und b): a) VARIO TOTAL NITROGEN HYDROX. LR, Set VARIO TOTAL NITROGEN HYDROX. LR, tubes VARIO TOTAL N PERSULFATE Reagent, b) VARIO TOTAL NITROGEN ACID LR/HR, Set VARIO TOTAL NITROGEN Reag. A VARIO TOTAL NITROGEN Reag. B VARIO TOTAL NITROGEN ACID LR/HR tubes Deionised water		■	■ ■ ■
Nitrogen, total HR	5 – 150 mg/l N	VARIO TOTAL NITROGEN HR, Set consists of a) und b): a) VARIO TOTAL NITROGEN HYDROX. HR, Set VARIO TOTAL NITROGEN HYDROX. HR, tubes VARIO TOTAL N PERSULFATE Reagent, b) VARIO TOTAL NITROGEN ACID LR/HR, Set VARIO TOTAL NITROGEN Reag. A VARIO TOTAL NITROGEN Reag. B VARIO TOTAL NITROGEN ACID LR/HR tubes Deionised water	■	■	■ ■ ■

instruments*

Method	Applications	Quantity	Code
4-(Dimethylamino)-benzaldehyde	Water, waste water, seawater	100 ml	53 12 00
Iron, total: 1, 10-phenantroline Iron, total: TPTZ	Water, waste water, seawater	100	53 05 60
	Water, waste water, seawater	100	53 05 50
		1 Set	53 60 10
Fe in Mo	Water, waste water	100	53 03 10
		100	53 03 20
		1 Set	53 50 90
PAN	Water, waste water	60 ml	
		100	
		60 ml	
Periodate oxidation	Water, waste water	1 Set	53 51 00
		100	
		100	
Mercaptoacetic acid	Water, waste water	1 Set	53 54 50
		100	
		100	
Mercaptoacetic acid	Water, waste water	1 Set	53 53 00
		100	
		100	
		100	
Mercaptoacetic acid	Water, waste water	1 Set	53 54 00
		100	
		100	
		100	
Chromotropic acid	Water, waste water	1 Set	53 55 80
		50	
		100 ml	
Persulfate digestion	Water, waste water	1 Set	53 55 50
		50	
		50	
		50	
		50	
		100 ml	
		1 Set	53 55 60
Persulfate digestion	Water, waste water	50	
		50	
		50	
		50	
		100 ml	



MSDS (Material Safety Data Sheets): www.lovibond.com

* HACH® is a registered trademark of Hach Company, Loveland, Colorado. The use of the HACH® trademark does not imply any affiliation with or approval by Hach Company regarding the formulation, testing or compatibility of these products for use in HACH® brand spectrophotometers or other devices or systems.

Reagents also compatible in Hach®

VARIO Powder Packs (PP) and Reagents

Test	Range	Reagent	Liquid Reagent	Tube Tests	Powder Pack
Nitrite LR	0 – 0.3 mg/l N	VARIO NITRI3, F10 VARIO NITRI3, F25			■ ■
Phosphate	0 – 2.5 mg/l PO ₄	VARIO PHOSPHATE RGT, F10			■
Phosphate, ortho	0.06 - 5 mg/l PO ₄	VARIO REACTIVE PHOSPHATE REAGENT SET consists of: VARIO PHOSPHATE DILUTION TUBE TEST VARIO PHOSPHATE RGT, F10 Deionised water	■	■	■
Phosphate, Acid hydrolyzable and total	acid hydrolyzable: 0.02 - 1.6 mg/l P Δ 0.06 - 5 mg/l PO ₄ total: 0.02 - 1.1 mg/l P Δ 0.06 - 3.5 mg/l PO ₄	VARIO TOTAL & ACID HYDROLYZABLE PHOSPHATE REAGENT SET consists of: VARIO PHOSPHATE ACID REAG. TUBE TEST Deionised water VARIO PHOSPHATE RGT, F10 VARIO SODIUM HYDROXID 1N VARIO SODIUM HYDROXID 1,54N VARIO POTASSIUM PERSULFATE	■ ■ ■	■	■ ■ ■
Phosphate, total	0.02 - 1.1 mg/l P Δ 0.06 - 3.5 mg/l PO ₄	VARIO TOTAL PHOSPHATE REAGENT SET consists of: VARIO PHOSPHATE ACID REAG. TUBE TEST VARIO PHOSPHATE RGT, F10 Deionised water VARIO SODIUM HYDROXID 1,54N VARIO POTASSIUM PERSULFATE	■ ■	■	■ ■
Phosphonates	0.02 - 125 mg/l PO ₄	VARIO PHOSPHONATE REAGENT SET consists of: VARIO Potassium Persulfate F10 VARIO PHOSPHATE RGT, F10			■ ■
Silica, LR	0 – 1.6 mg/l SiO ₂	VARIO SILICA Reagent LR, Set F10 consists of: VARIO LR SILICA AMINO ACID F VARIO SILICA CITRIC ACID VARIO MOLYBDATE 3 Reagent solution	■		■ ■
Silica, HR	0 – 100 mg/l SiO ₂	VARIO SILICA Reagent HR, Set F10 consists of: VARIO SILICA HR MOLYBDATE, F10 VARIO SILICA HR ACID RGT, F10 VARIO SILICA CITRIC ACID, F10			■ ■ ■
Silica, UHR	0 – 200 mg/l SiO ₂	VARIO SILICA Reagent HR, Set F25 consists of: VARIO SILICA HR MOLYBDATE, F25 VARIO SILICA HR ACID RGT, F25 VARIO SILICA HR CITRIC ACID, F25			■ ■ ■
Sulphate	0 – 70 mg/l SO ₄	VARIO Sulpha 4, F10 VARIO Sulpha 4, F25			■ ■
Triazoles	1 - 16 mg/l	VARIO Triazole Rgt F25			■

devices*

Method	Applications	Quantity	Code
Diazotiation	Water, waste water	100 100	53 09 80 53 09 70
Phosphomolybdic acid/ Ascorbic acid	Water, waste water, seawater	100	53 15 50
		1 Set	53 52 00
Phosphomolybdic acid/ Ascorbic acid	Water, seawater	50 50 100 ml	
		1 Set	53 52 50
Phosphomolybdic acid/ Ascorbic acid	Water, seawater	50 50 100 ml 100 ml 100 ml 50	
		1 Set	53 52 10
Phosphomolybdic acid/ Ascorbic acid	Water, seawater	50 50 100 ml 100 ml 50	
		1 Set	53 52 20
Persulfate UV-Oxidation	Water	100 200	
		1 Set	53 56 90
Heteropoly blue	Water, seawater	100 200 2 x 50 ml	
		1 Set	53 57 00
Silicomolybdate	Water, seawater	100 100 100	
		1 Set	53 59 00
Silicomolybdate	Water, seawater	100 100 100	
USEPA accepted for waste water analysis	Water, waste water, seawater	100 100	53 21 60 53 21 50
Catalyzed UV Digestion	Water	100	53 22 00



MSDS (Material Safety Data Sheets): www.lovibond.com

* HACH® is a registered trademark of Hach Company, Loveland, Colorado. The use of the HACH® trademark does not imply any affiliation with or approval by Hach Company regarding the formulation, testing or compatibility of these products for use in HACH® brand spectrophotometers or other devices or systems.