# Lovibond® Colour Measurement Tintometer® Group



## CIE L\*a\*b\*



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### CIE L\*a\*b\*

A single glass conformance filter for checking the accuracy of colour instruments in CIE  $L^*a^*b^*$  colour space.

Bestell-Nr.: 130501

#### Technische Daten

Glass Standard Values Näherungswerte - L\* 92a - 3.4 b\*26

Filters

## Lieferumfang

CIE filter

Certificate of conformity