

Lovibond® Colour Measurement Tintometer® Group



Conjunto de filtros Número de cor ASTM (PFXi195)

Artigo No: 139810



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Conjunto de filtros Número de cor ASTM (PFXi195)

A set of three conformance filters to check the PFXi instruments in ASTM D1500 scale.

Dados Técnicos

Glass Standard Values	Valores aproximados: -
	0.5
	3.5
	5.0
Filters	3

Texto do Escopo de Entrega

- 0.5
- 3.5
- 5.0
- Certificate of Conformity