

Lovibond® Colour Measurement Tintometer® Group



Conjunto de filtros Número de cor Gardner (PFXi995/950/880)

Artigo No: 139890



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Conjunto de filtros Número de cor Gardner (PFXi995/950/880)

A set of four glass conformance filters for checking PFXi instruments in Gardner scale.

Dados Técnicos

Glass Standard Values	Valores aproximados: -
	2
	8
	12
	17
Filters	4

Texto do Escopo de Entrega

- Gardner 2 filter
- Gardner 8 filter
- Gardner 12 filter
- Gardner 17 filter
- Certificate of conformity