Lovibond® Colour Measurement Tintometer® Group



Lovibond® Kırmızı, Sarı, Mavi (Model F, E, D)

Ürün No: 186000



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lovibond® Kırmızı, Sarı, Mavi (Model F, E, D)

A set of five glass conformance filters for checking a Model F, E and D in Lovibond RYBN scale.

Teknik Veri

Glass Standard Valu- Approximate Values 1.2R 0.9B 1.1R 4.3Y 1.4B - 2.6R 11.5Y 2.1B 3.2R 15.5Y 2.9B1.0Y 0.5B

Filters

Teslimat Kapsam Metni

- Lovibond RYBN 1.2R 0.9B filter
- Lovibond RYBN 1.1R 4.3Y 1.4B filter
- Lovibond RYBN 2.6R 11.5Y 2.1B filter
- Lovibond RYBN 3.2R 15.5Y 2.9B filter
- Lovibond RYBN 1.0Y 0.5B filter
- Certificate of conformity