

**Nitrite VHR L****M271****25 - 2500 mg/L NO<sub>2</sub><sup>-</sup>****Ferrous Sulfate Method**

## Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range
MD 600, MD 610, MD 640	ø 24 mm	580 nm	25 - 2500 mg/L NO <sub>2</sub> <sup>-</sup>
XD 7000, XD 7500	ø 24 mm	585 nm	25 - 2500 mg/L NO <sub>2</sub> <sup>-</sup>

## Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Nitrite VHR L, 500 ml	500 mL	471170
Nitrite VHR L, 500 ml, Set	500 mL	471160

The following accessories are required.

Accessories	Packaging Unit	Part Number
Pipette, 1000 µl	1 pc.	365045
Pipette tips, 0,1-1 ml (white), 1000 pc.	1 pc.	419073

## Application List

- Cooling Water




## Determination of Nitrite VHR L


Select the method on the device.



Place **10 mL Nitrite VHR L solution** in the sample cuvette.




Close vial(s).




Place **sample vial** in the sample chamber. Pay attention to the positioning.




Press the **ZERO** button.




Remove the vial from the sample chamber.




Add **1 mL sample**.



Close vial(s).



Invert several times to mix the contents (1-2 times).



Place **sample vial** in the sample chamber. Pay attention to the positioning.

# Test

Press the **TEST** (XD:  
**START**)button.

The result in mg/L Nitrite appears on the display.

## Chemical Method

Ferrous Sulfate Method

### Calibration function for 3rd-party photometers

$$\text{Conc.} = a + b \cdot \text{Abs} + c \cdot \text{Abs}^2 + d \cdot \text{Abs}^3 + e \cdot \text{Abs}^4 + f \cdot \text{Abs}^5$$

	<b>ø 24 mm</b>	<b>□ 10 mm</b>
a	1.45432•10 <sup>+0</sup>	1.45432•10 <sup>+1</sup>
b	1.22994•10 <sup>+3</sup>	2.64437•10 <sup>+3</sup>
c		
d		
e		
f		

### Method Validation

<b>Limit of Detection</b>	8.77 mg/L
<b>Limit of Quantification</b>	26.31 mg/L
<b>End of Measuring Range</b>	2500 mg/L
<b>Sensitivity</b>	1235.02 mg/L / Abs
<b>Confidence Intervall</b>	13.11 mg/L
<b>Standard Deviation</b>	5.42 mg/L
<b>Variation Coefficient</b>	0.43 %