

M284

ΤN	HR	2	TT

5 - 140 mg/L N^{b) i)}

2,6-Dimethylphenole

Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range
SpectroDirect, XD 7000, XD 7500	ø 16 mm	340 nm	5 - 140 mg/L N ^{b) i)}

Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Total Nitrogen DMP HR / 25	1 pc.	2423570
Total Nitrogen	1 pc.	2420703

The following accessories are required.

Accessories	Packaging Unit	Part Number
Thermoreactor RD 125	1 pc.	2418940

Application List

- · Waste Water Treatment
- Drinking Water Treatment
- · Raw Water Treatment

Notes

1. This test determines the inorganic compounds Ammonia, Nitrate and Nitrite, as well as organic compounds like amino acid, urea, complexing agents etc.



TN HR 2 TT / M284




Digestion





Put **4.5 mL deionised water** in the digestion vial.



Add a level measuring scoop No. 8 (black) Digestion Reagent .




Close vial(s).

digestion vial.



Invert several times to mix the contents.




Seal the vials in the preheated thermoreactor for 60 minutes at 100 °C.



Remove the vial from the thermoreactor. (Note: vial will be hot!)




Invert several times to mix the contents.




Allow the vial(s) to cool to room temperature.



TN HR 2 TT / M284







Close vial(s).



Invert several times to mix the contents.



Determination of Nitrogen, total HR with Vial Test


Select the method on the device.

For testing of Nitrogen, total HR with tube test, carry out the described digestion.

For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500








Place the supplied Zero vial (red sticker) in the sample chamber. • Pay attention to the positioning.

Place the supplied Zero vial Press the **ZERO** button.


Remove **vial** from the sample chamber.

For devices that require no ZERO measurement, start here.



Open a digestion vial.





Fill sample vial with **0.5 mL** Close vial(s). prepared, digested sample.



Carefully invert several times to mix the contents. **Note: Will get hot!**





Add 0.2 mL Nitrate-111.


Close vial(s).




TN HR 2 TT / M284




Invert several times to mix the contents.



Place **sample vial** in the sample chamber. • Pay attention to the positioning.



Press the **TEST** (XD: **START**)button.



Wait for 15 minute(s) reaction time.

Once the reaction period is finished, the measurement takes place automatically.

The result in mg/L Nitrogen appears on the display.



Analyses

The following table identifies the output values can be converted into other citation forms.

Unit	Cite form	Scale Factor
mg/l	Ν	1
mg/l	NH₄	1.288
mg/l	NH ₃	1.2158

Chemical Method

2,6-Dimethylphenole

Appendix

Calibration function for 3rd-party photometers

Conc. = $a + b \cdot Abs + c \cdot Abs^2 + d \cdot Abs^3 + e \cdot Abs^4 + f \cdot Abs^5$

	ø 16 mm
а	-9.36243 • 10 ⁻¹
b	2.51666 • 10 ^{∗1}
c	
d	
e	
f	

Interferences

Persistant Interferences

• Nitrogen compounds which are hardly to oxidise, as may be found in industrial sewage, are not digested or only partially.

Bibliography

1. ISO 23697-1, Water quality — Determination of total bound nitrogen (ST-TNb) in water using small-scale sealed tubes — Part 1: Dimethylphenol colour reaction

According to

US EPA 40 CFR 141

Derived from

EN ISO 11905-1



 $^{\rm b)}$ Reactor is necessary for COD (150 °C), TOC (120 °C) and total -chromium, - phosphate, -nitrogen, (100 °C) | $^{\rm b}$ high range by dilution