

Lovibond® Colour Measurement Tintometer® Group



Betacaroteno

Referencia No: 169520



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Betacaroteno

A glass conformance filter to check the Model Fx range of instruments accuracy in measuring beta Carotene.

Datos técnicos

Glass Standard Values – 45 ppm at 10 mm path length

Filters 1

Volumen de suministro

- 45ppm filter
- Certificate of Conformity