

# Lovibond® Colour Measurement Tintometer® Group



## Lot de filtres, indice Hazen (uniquement PFX/ PFXi 880/950/995)

Code: 135159



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Lot de filtres, indice Hazen (uniquement PFX/PFXi 880/950/995)

Glass conformance filters for checking Pt-Co scale on EC 2000 PT-Co.

#### Fiche technique

<b>Glass Standard Values</b>	Valeurs approximatives :
	5,
	20
	50
	100
	300
<b>Filters</b>	5

### Contenu de livraison

- 5, 20, 50, 100, 300 filters
- Certificate of Conformity