# Lovibond<sup>®</sup> Colour Measurement **Tintometer<sup>®</sup> Group**



## Bêta-carotène



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

#### Code: 169520

#### Bêta-carotène

A glass conformance filter to check the Model Fx range of instruments accuracy in measuring beta Carotene.

#### Fiche technique

Glass Standard Va-45 ppm avec un trajet optique de 10 mm lues 1

Filters

### Contenu de livraison

45ppm filter

Certificate of Conformity