

Lovibond® Colour Measurement Tintometer® Group



Set di filtri EP Marrone/Giallo

N. ordine: 139440



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Set di filtri EP Marrone/Giallo

A set of two glass conformance filters for checking PFXi instruments in European Pharmacopoeia Brown/Yellow scale.

Dati tecnici

Glass Standard Values	- BY3
	- BY5

Filters	2
----------------	---

Fornitura

- EP Brown/Yellow 3 filter
- EP Brown/Yellow 5 filter
- Certificate of Conformity