

Lovibond® Colour Measurement Tintometer® Group



Chlorofyl A (PFXi 995/950/880)

Bestelnr.: 139530



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Chlorofyl A (PFXi 995/950/880)

A glass conformance filter to check the PFXi (PFXi 995/950/880) range of instruments accuracy in measuring Chlorophyll A.

Technische gegevens

Glass Standard Values - 0.19 ppm at 10 mm path length

Filters 1

Leveromvang

- 0.19 ppm filter
- Certificate of Conformity