

Lovibond® Colour Measurement Tintometer® Group



Número de cor ASTM

Artigo No: 341000



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Número de cor ASTM

A set of three glass filters to check the AF650 Petroleum Oils Comparator in ASTM D1500 scale.

Dados Técnicos

Glass Standard Values	Valores aproximados: - 0.5, 3.5, 5.0
Filters	3

Texto do Escopo de Entrega

- 0.5
- 3.5
- 5.0
- Certificate of conformity