

Lovibond® Colour Measurement Tintometer® Group



EBC

Номер заказа: 131044



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

EBC

A single glass conformance filter for checking PFXi instruments in EBC scale. This filter was previously provided as part of the accessories kit for the PFX195/4

Технические данные

Glass Standard Values	Approximate Value – 7.5
Filters	1

Объем поставки

- EBC 7.5 filter
- Certificate of Conformity