

# Lovibond® Colour Measurement Tintometer® Group



## Набор фильтров для шкалы IP

Номер заказа: 137660



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Набор фильтров для шкалы IP

A set of four glass conformance filters for checking the PFXi instruments in IP (method B) units

#### Технические данные

Glass Standard Values	Approximate values
	- 0.75
	- 1.25
	- 2.0
	- 3.5

Filters 4

### Объем поставки

- IP 0.75 filter
- IP 1.25 filter
- IP 2.0 filter
- IP 3.5 filter
- Certificate of conformity