# Lovibond® Colour Measurement Tintometer® Group



## CIE L\*a\*b\* (PFXi)



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

#### CIE L\*a\*b\* (PFXi)

A single glass conformance filter for checking the accuracy of PFXi instruments in CIE xyY colour space.

#### Teknik Veri

Glass Standard Valu- Approximate values es - L\* 94 a\* - 2.9 b\*16

Ürün No: 130500

Filters

1

### Teslimat Kapsam Metni

- CIE filter
- Certificate of conformity