

Lovibond® Colour Measurement Tintometer® Group



Klett

Ürün No: 139713



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Klett

A single glass conformance filter for checking Klett scale on PFXi instruments.

Teknik Veri

Glass Standard Values	Approximate Value
	– 162

Filters	1
---------	---

Teslimat Kapsam Metni

- Klett 162 filter
- Certificate of conformity