# Lovibond<sup>®</sup> Colour Measurement

## Tintometer<sup>®</sup> Group



### Chlorophyll A (Fx)



Conformance filters are assigned a nominal value and the company always endeavors to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

#### Part Number: 169530

#### Chlorophyll A (Fx)

A glass conformance filter to check the PFXi range of instruments accuracy in measuring Chlorophyll A.

#### **Technical Data**

Glass Standard Values	-	0.19 ppm at 10 mm path length
Filters	1	

### **Delivery Scope**

0.19 ppm filter Certificate of Conformity