

Lovibond® Colour Measurement Tintometer® Group



AOCS

Referencia No: 136980



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

AOCS

A single glass filter for testing the PFXi 880, 950 or 995 in AOCS-Tintometer scale.

Datos técnicos

Glass Standard Values	Näherungswert – 0,5 R, 1,0 Y
Filters	1

Volumen de suministro

- 0.5R 1.0Y
- Certificate of Conformity