

# Lovibond® Colour Measurement Tintometer® Group



## Clorofila A (Fx)

Artigo No: 169530



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Clorofila A (Fx)

A glass conformance filter to check the Model Fx range of instruments accuracy in measuring Chlorophyll A.

### Dados Técnicos

<b>Glass Standard Values</b>	Extensão de percurso de 0,19 ppm a 10 mm
<b>Filters</b>	1

## Texto do Escopo de Entrega

- 0.19 ppm filter
- Certificate of Conformity