

# Lovibond® Colour Measurement Tintometer® Group



## EBC

Artigo No: 344000



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### EBC

A set of four glass conformance filters for checking the 3000 Comparator in EBC scale

### Dados Técnicos

<b>Glass Standard Values</b>	Valores aproximados: - 4, 9.5, 15, 25
<b>Filters</b>	4

## Texto do Escopo de Entrega

- EBC 4 filter
- EBC 9.5 filter
- EBC 15 filter
- EBC 25 filter
- Certificate of Conformity